
寄件者： RYAN
收件者： "coco_huang"; "Nick"
副本： "Killy_Lee"; "LaslieChen 陳志偉"; "asli_cheng"; "JoanneHuang 黃瓊玉-生產課(ENG/PCD)"
主旨： RE: Packaging Label of serenity 1A42550000 & 1A42550001Carton label 線上會簽 (PM急需要BMP檔給瓊玉)
日期： 2025年4月29日 上午 11:19:26
附件： image008.png
image012.png

Hi Coco,

請幫我修改如下圖, 立即導入. 謝謝



Best Regards,
Ryan Wang

DIVA Laboratories , Ltd.
Tel: (02)2226-8631 EXT: 223

From: coco_huang <coco_huang@diva.com.tw>
Sent: Monday, April 28, 2025 3:42 PM
To: 'Nick' <nick_lin@diva.com.tw>; 'RYAN' <ryan_wang@diva.com.tw>
Cc: 'Killy_Lee' <killy_lee@diva.com.tw>; 'LaslieChen 陳志偉' <lasliechen@diva.com.tw>; 'asli_cheng' <asli_cheng@diva.com.tw>; 'JoanneHuang 黃瓊玉-生產課(ENG/PCD)' <joannehuang@diva.com.tw>
Subject: RE: Packaging Label of serenity 1A42550000 & 1A42550001Carton label 線上會簽 (PM急需要BMP檔給瓊玉)
Importance: High

Dear瓊玉:

PN碼	客戶機種名稱	Carton label
1A42550000	BEZEL USER INTERFACE – T	8MD672LB02B
1A42550001	BEZEL USER INTERFACE – S	8MD672LB12B

剛剛Ryan來電通知我，**請我先將以上2個機種的已變更過後的Carton label BMP檔給您**。
如有問題再請您詢問Ryan，謝謝。

Best Wishes
coco_huang
DIVA Laboratories, Ltd. | Mechanical Engineering Dep.
T:+886-2-2226-8631#509 | E: coco_huang@diva.com.tw
F:+886-2-2226-2423 | W: www.diva.com.tw

A: 9F., No.351, Sec. 2, Zhongshan Road, Zhonghe District , New Taipei City 235602, Taiwan R.O.C.
A: 235602新北市中和區中山路二段351號9樓

From: Nick [mailto:nick_lin@diva.com.tw]
Sent: Monday, April 28, 2025 2:56 PM
To: 'RYAN' <ryan_wang@diva.com.tw>; 'coco_huang' <coco_huang@diva.com.tw>
Cc: 'Killy_Lee' <killy_lee@diva.com.tw>; 'LaslieChen 陳志偉' <lasliechen@diva.com.tw>; 'asli_cheng' <asli_cheng@diva.com.tw>; 'JoanneHuang 黃瓊玉-生產課(ENG/PCD)' <joannehuang@diva.com.tw>
Subject: RE: Packaging Label of serenity 1A42550000 & 1A42550001Carton label 線上會簽

Dear 瓊玉

換你

謝謝

Warm Regards
Nick_lin

DIVA Laboratories, Ltd. | PE Dep.
T:+886-2-2226-8631#724 | E: nick_lin@diva.com.tw
F:+886-2-2226-2423 | W: www.diva.com.tw
A: 9F., No.351, Sec. 2, Zhongshan Road, Zhonghe District , New Taipei City 235602, Taiwan R.O.C.

From: RYAN [mailto:ryan_wang@diva.com.tw]
Sent: Monday, April 28, 2025 2:04 PM
To: 'coco_huang' <coco_huang@diva.com.tw>
Cc: 'Killy_Lee' <killy_lee@diva.com.tw>; 'LaslieChen 陳志偉' <lasliechen@diva.com.tw>; 'asli_cheng' <asli_cheng@diva.com.tw>; 'Nick_Lin 林益銓' <nick_lin@diva.com.tw>; 'JoanneHuang 黃瓊玉-生產課(ENG/PCD)' <joannehuang@diva.com.tw>
Subject: RE: Packaging Label of serenity 1A42550000 & 1A42550001Carton label 線上會簽

Hi Nick,

換你. 謝謝

Best Regards,
Ryan Wang

DIVA Laboratories , Ltd.
Tel: (02)2226-8631 EXT: 223

From: coco_huang <coco_huang@diva.com.tw>
Sent: Monday, April 28, 2025 1:45 PM
To: 'RYAN' <ryan_wang@diva.com.tw>
Cc: 'Killy_Lee' <killy_lee@diva.com.tw>; 'LaslieChen 陳志偉' <lasliechen@diva.com.tw>; 'asli_cheng' <asli_cheng@diva.com.tw>; 'Nick_Lin 林益銓' <nick_lin@diva.com.tw>; JoanneHuang 黃瓊玉-生產課(ENG/PCD) <joannehuang@diva.com.tw>
Subject: RE: Packaging Label of serenity 1A42550000 & 1A42550001Carton label 線上會簽
Importance: High

Dear all
再麻煩請會簽以下2個機種的Carton label 會簽，謝謝。

PN碼	客戶機種名稱
1A42550000	BEZEL USER INTERFACE – T
1A42550001	BEZEL USER INTERFACE – S

會簽順序:
Ryan → Nick → Joanne

Best Wishes
coco_huang
DIVA Laboratories, Ltd. | Mechanical Engineering Dep.
T:+886-2-2226-8631#509 | E: coco_huang@diva.com.tw
F:+886-2-2226-2423 | W: www.diva.com.tw
A: 9F., No.351, Sec. 2, Zhongshan Road, Zhonghe District , New Taipei City 235602, Taiwan R.O.C.
A: 235602新北市中和區中山路二段351號9樓

From: RYAN [mailto:ryan_wang@diva.com.tw]
Sent: Monday, April 28, 2025 1:32 PM
To: 'coco_huang' <coco_huang@diva.com.tw>
Cc: 'Killy_Lee' <killy_lee@diva.com.tw>; 'LaslieChen 陳志偉' <lasliechen@diva.com.tw>; 'asli_cheng' <asli_cheng@diva.com.tw>; 'Nick_Lin 林益銓' <nick_lin@diva.com.tw>
Subject: RE: Packaging Label of serenity 1A42550000 & 1A42550001Carton label修改並料號進階

Hi Coco,

Carton label 客戶確認 OK. 謝謝

Mon 4/28/2025 12:37 PM

Jiang, Xinkai <Xinkai.Jiang@gehealthcare.com>

RE: Packaging Label of serenity

收件者 RYAN

副本 'Laslie'; 'asli_cheng'; 'Nick_Lin 林益銓'; 'coco_huang'

Hi Ryan,


No issues, OK for the lable.

From: RYAN <ryan_wang@diva.com.tw>
Sent: Friday, April 25, 2025 12:03 PM
To: Jiang, Xinkai <Xinkai.Jiang@gehealthcare.com>
Cc: 'Laslie' <lasliechen@diva.com.tw>; 'asli_cheng' <asli_cheng@diva.com.tw>; 'Nick_Lin 林益銓' <nick_lin@diva.com.tw>; 'coco_huang' <coco_huang@diva.com.tw>
Subject: RE: Packaging Label of serenity

WARNING: This email originated from outside of GE HealthCare. Please validate t

Hi Xinkai,

Update carton label design as below. Please check and let us know if there is any question. Thank you



Best Regards,
Ryan Wang

Best Regards,
Ryan Wang

DIVA Laboratories , Ltd.
Tel: (02)2226-8631 EXT: 223

From: coco_huang <coco_huang@diva.com.tw>
Sent: Thursday, April 24, 2025 10:37 AM
To: 'Killy_Lee' <killy_lee@diva.com.tw>; 'Ryan_Wang 王志維' <ryan_wang@diva.com.tw>
Subject: RE: Packaging Label of serenity 1A42550000 & 1A42550001Carton label修改並料號進階

Dear Ryan:

PN碼	客戶機種名稱
1A42550000	BEZEL USER INTERFACE – T (3.7)(Black)
1A42550001	BEZEL USER INTERFACE – S (3.7)(White)

修改後的Carton label如附件，再請看一下，是否ok謝謝~
如沒問題再請回覆，後續才可以做**會簽**跟**EC的作業**。


Best Wishes

coco_huang
DIVA Laboratories, Ltd. | **Mechanical Engineering Dep.**
T:+886-2-2226-8631#509 | **E: coco_huang@diva.com.tw**

F:+886-2-2226-2423 | W: www.diva.com.tw
A: 9F., No.351, Sec. 2, Zhongshan Road, Zhonghe District , New Taipei City 235602, Taiwan R.O.C.
A: 235602新北市中和區中山路二段351號9樓

From: Killy_Lee [mailto:killy_lee@diva.com.tw]
Sent: Wednesday, April 23, 2025 4:36 PM
To: coco_huang <coco_huang@diva.com.tw>; Ryan_Wang 王志維 <ryan_wang@diva.com.tw>
Subject: FW: Packaging Label of serenity

Hi Coco
請協助修正這個檔案
依照 GEHC SQE 的回覆, 3.7” carton label 修改如下. 8MD672LB12A & 8MD672LB02A請同步修改. 謝謝




謝謝您
Killy lee (美嬿)

From: RYAN [mailto:ryan_wang@diva.com.tw]
Sent: Wednesday, April 23, 2025 4:28 PM
To: 'Killy_Lee' <killy_lee@diva.com.tw>
Cc: 'Nick_Lin 林益銓' <nick_lin@diva.com.tw>; 'lasliechen' <lasliechen@diva.com.tw>; 'asli_cheng' <asli_cheng@diva.com.tw>; 'Oliver_Chen 陳冠華' <oliver_chen@diva.com.tw>
Subject: RE: Packaging Label of serenity

Hi Killy,

依照 GEHC SQE 的回覆, 3.7” carton label 修改如下. 8MD672LB12A & 8MD672LB02A請同步修改. 謝謝



Best Regards,
Ryan Wang

DIVA Laboratories , Ltd.
Tel: (02)2226-8631 EXT: 223

From: Jiang, Xinkai <Xinkai.Jiang@gehealthcare.com>
Sent: Wednesday, April 23, 2025 10:15 AM
To: Ryan_Wang <ryan_wang@diva.com.tw>
Subject: Packaging Label of serenity

Hi Ryan,

Per document 46-274914, the label is lack of some information.

1. No part Revision
2. No PO number
3. No country of origin
4. No vender code

BTW, DIVA's vender code is 821872

Please modify the Label and take a picture to let me check ,thanks.

Thank you,
Best Regards

蒋心 （Xinkai Jiang）
Supplier Quality Development
GE HealthCare
M +86 180 2118 9125
Xinkai.jiang@gehealthcare.com
gehealthcare.com